

# Multicenter Clinical Evaluation of a Real-Time PCR Assay for *Bordetella pertussis*

T. S. Uphoff<sup>1</sup>, J. Dien Bard<sup>2</sup>, G. Denys<sup>3</sup>, M. Oethinger<sup>4</sup>, W. Sinclair<sup>5</sup>, S. Young<sup>6</sup>

<sup>1</sup>Marshfield Labs, Marshfield, WI; <sup>2</sup>Children's Hospital and University of Southern California School of Medicine, Los Angeles, CA; <sup>3</sup>Indiana University School of Medicine, Indianapolis, IN; <sup>4</sup>Providence **St. Joseph's Health**, Oregon, Portland, OR; <sup>5</sup>Intermountain Healthcare, Salt Lake City, UT; <sup>6</sup>TriCore Laboratories, Albuquerque, NM

## Background

Pertussis is a contagious respiratory disease caused by *Bordetella (B.) pertussis* that can lead to life-threatening complications in young children. Early and accurate diagnosis is critical for positive patient outcomes and to aid public health officials in controlling spread during outbreaks. The Great Basin Bordetella Direct Test is a qualitative *in vitro* diagnostic test for the detection of *B. pertussis* DNA from nasopharyngeal (NP) swab specimens obtained from patients suspected of having a respiratory tract infection from *B. pertussis*. The assay was performed on the PA500 Portrait™ Analyzer and utilizes PCR amplification of the insertion sequence IS481.

## Methods

The performance of the Bordetella Direct Test was evaluated on 1037 NP samples, 915 prospectively collected and 122 retrospective frozen NP samples at 5 US clinical sites from August, 2016 to January, 2017. Results generated by the Bordetella Direct Test were compared to those obtained with the Nanosphere Verigene® RP Flex Test. Discordant samples were adjudicated by testing with the Quidel Amplivue *B. pertussis*, an FDA approved molecular assay. Reproducibility studies were performed at 3 sites with 2 operators at each site. A 3 sample panel consisting of a moderate positive sample (1.9x the limit of detection {LoD}) a high positive sample (3.8x LoD) and a negative sample was tested in triplicate over 5 non-consecutive days by each operator (n=270).

## Results

In the combined evaluation of the Bordetella Direct Test (prospective and retrospective samples), the Positive Percent Agreement (PPA) for *B. pertussis* was 92.5% (74/80; 95% confidence interval [CI], 84.6% to 96.5%). The Negative Percent Agreement (NPA) for *B. pertussis* was 99.6% (953/957; 95% CI, 98.9% to 99.8%). Six samples were positive by the RP Flex Test and negative by the Bordetella Direct Test, 4 of the 6 were confirmed positive using the Amplivue assay. Conversely, 4 samples were Bordetella Direct Test positive and RP Flex negative and 2 of these were confirmed positive using the Amplivue assay.

## Results Continued

**Table 1.** Summary of Clinical Study Results

| Specimen                    | n                         | % Agreement (95% CI) |                                   |                                     |
|-----------------------------|---------------------------|----------------------|-----------------------------------|-------------------------------------|
|                             |                           | Positive             | Negative                          |                                     |
| <i>Bordetella pertussis</i> | Fresh                     | 915                  | 85.7%<br>(65.4% - 95.0%)<br>18/21 | 99.6%<br>(98.9% - 99.8%)<br>890/894 |
|                             | Frozen                    | 122                  | 94.6%<br>(86.1% - 98.3%)<br>56/59 | 100.0%<br>(94.3% - 100.0%)<br>63/63 |
|                             | Combined (Fresh + Frozen) | 1037                 | 92.5%<br>(84.6% - 96.5%)<br>74/80 | 99.6%<br>(98.9% - 99.8%)<br>953/957 |

**Table 2.** Summary Table – *B. pertussis* False Negative Discrepant Results – Prospective Study

| # | Sample ID | Bordetella Direct Test<br><i>B. pertussis</i><br>Result | Verigene® RP Flex<br><i>B. pertussis</i><br>Result | Amplivue<br><i>B. pertussis</i><br>Result |
|---|-----------|---|--|---|
| 1 | TRC-21    | Not Detected  | Detected   | Positive                                  |
| 2 | MCL-171   | Not Detected  | Detected   | Positive                                  |
| 3 | MCL-182   | Not Detected  | Detected   | Negative                                  |

**Table 3.** Summary Table – *B. pertussis* False Positive Discrepant Results – Prospective Study

| # | Sample ID | Bordetella Direct Test<br><i>B. pertussis</i><br>Result | Verigene® RP Flex<br><i>B. pertussis</i><br>Result | Amplivue<br><i>B. pertussis</i><br>Result |
|---|-----------|---|--|---|
| 1 | MCL-5     | Detected  | Not Detected                                       | Negative                                  |
| 2 | MCL-13    | Detected  | Not Detected                                       | Positive                                  |
| 3 | MCL-273   | Detected  | Not Detected                                       | Negative                                  |
| 4 | MCL-477   | Detected  | Not Detected                                       | Positive                                  |

## Results Continued

**Table 4.** Summary Table – *B. pertussis* Discrepant Results – Frozen Archived Sample

| # | Sample ID | Bordetella Direct Test<br><i>B. pertussis</i><br>Result | Verigene® RP Flex<br><i>B. pertussis</i><br>Result | Amplivue<br><i>B. pertussis</i><br>Result |
|---|-----------|---|--|---|
| 1 | FR-046    | Not Detected  | Detected   | Positive                                  |
| 2 | FR-064    | Not Detected  | Detected   | Negative                                  |
| 3 | FR-085    | Not Detected  | Detected   | Positive                                  |

% positive agreement after discrepancy analysis was 95% (76/80) and % negative agreement after discrepancy analysis was 99.8% (955/957)

**Table 5.** Reproducibility Study Panel

| Panel | Sample Type       | Sample Composition               | Concentration (CFU/mL), LoD  |
|-------|-------------------|----------------------------------|------------------------------|
| RP-01 | Low Positive      | Bordetella pertussis (ATCC 9797) | 3.1 x 10 <sup>3</sup> , 1.9X |
| RP-02 | Moderate Positive | Bordetella pertussis (ATCC 9797) | 6.2 x 10 <sup>3</sup> , 3.8X |
| RP-05 | Negative          | Negative                         | N/A                          |

There were no discordant results in the reproducibility study, all 1.9x, 3.8x LoD and negative samples gave the expected results in all testing events (270/270).

## Conclusions

The Great Basin Bordetella Direct Test is a sensitive and specific diagnostic assay for detection of *B. pertussis* in NP swabs from subjects with suspected *B. pertussis* infection.

\*We thank Great Basin Scientific for providing the financial support for this study.